

# SkillGuide

Important Product Information



## SkillGuide English

---

### Environmental Specifications

#### Temperature and Humidity

Operating temperature range: 0 °C to +40 °C, Humidity <95% R.H. Non-condensing.  
Storage and shipment temperature range: -15 °C to +50 °C

### Resistance to ingress of water and dust

Ingress of solid foreign objects - class IP3X in accordance with IEC60529.  
Ingress of water - class IPX2 in accordance with IEC60529.

### Water and moisture

Do not operate this device under or near water. Do not operate the product with wet hands.

### Cleaning

Disconnect SkillGuide from the mainline before cleaning or polishing. Do not use liquid cleaners or aerosol cleaners. Use a lint-free cloth lightly moistened with water for cleaning the exterior.

### Servicing

Do not attempt to disassemble or service SkillGuide. Disassembly and service must only be done by qualified service personnel.

### Limited warranty

Please refer to the Laerdal Global Warranty statement, and for more information see: [www.laerdal.com](http://www.laerdal.com)

Product specifications are subject to change without notice.



**WEEE**

### Waste Electrical and Electronic Equipment

This appliance is marked according to the European directive 2012/19/EU on Waste Electrical and Electronic Equipment (WEEE). By ensuring this product is disposed of correctly, you will help prevent potential negative consequences for the environment and human health, which could otherwise be caused by inappropriate waste handling of this product. The symbol on the product, or on the documents accompanying the product, indicates that this appliance may not be treated as household waste. Instead it shall be handed over to the applicable collection point for the recycling of electrical and electronic equipment. Disposal must be carried out in accordance with local environmental regulations for waste disposal. For more detailed information about treatment, recovery and recycling of this product, please contact your local city office, your household waste disposal service or Laerdal representative.

**Federal Communications Commission (FCC) and Industry Canada (IC) Statements**

This device complies with part 15 of the FCC Rules and RSS-210 of IC rules. Operation is subject to the following two conditions:

1. This device may not cause harmful interference, and
2. This device must accept any interference received, including interference that may cause undesired operation.

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Consult the dealer or an experienced radio/TV technician for help.

Caution: Changes or modifications not expressly approved by Laerdal Medical could void the user's authority to operate the equipment.

The term "IC" before the equipment certification number only signifies that the Industry Canada technical specifications were met.

Mise en garde: Tout changement ou toute modification n'ayant pas fait l'objet d'une approbation expresse de Laerdal Medical peut annuler le droit dont dispose l'utilisateur de se servir de l'équipement.

Le terme « IC » qui précède le numéro d'agrément de l'équipement signifie uniquement que les caractéristiques techniques spécifiées par Industrie Canada sont respectées.

This device contains licence-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's licence-exempt RSS(s). Operation is subject to the following two conditions:

1. This device may not cause interference.
2. This device must accept any interference, including interference that may cause undesired operation of the device.

L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

1. L'appareil ne doit pas produire de brouillage;
2. L'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.



Laerdal Medical AS hereby declares that when carrying the CE-mark, the product is in conformity with Essential Requirement of the Council Directive 2014/30/EU relating to Electromagnetic Compatibility (EMC); and Council Directive 2011/65/EU relating to Restriction of the use of certain Hazardous Substances (RoHS).



UKCA (U.K. Conformity Assessed): This product complies with the requirements of U.K. legislation of Electromagnetic Compatibility Regulations (S.I. 2016/1101).

UKCA (U.K. Conformity Assessed): This product complies with the requirements of U.K. legislation of Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations (RoHS) Regulations ((S.I. 2012/3032).

© 2022 Laerdal Medical AS. All rights reserved.



Manufacturer:

Laerdal Medical AS

P.O. Box 377

Tanke Svilandsgate 30, 4002 Stavanger, Norway

T: (+47) 51 51 17 00

20-06567 Rev D

[www.laerdal.com](http://www.laerdal.com)



**Laerdal**  
helping save lives