

Manufacturer: Beurer GmbH (see address in footer)

Product category: Blood glucose monitor

Product type: GL 44

The product specified above is in conformity with the following specifications.

98/79/EC In vitro medical device directive (IVDD)

EDMA code and name: 21-06-01 Blood glucose meter

Classification: Annex II/list B

Conformity assessment procedure: Annex IV section 3

The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:

Certificate no. and validity: D1311700053, valid to 2024-12-17

EN 61010-1:2010  
EN 61010-2-101:2017  
EN 61326-1:2013  
EN 61326-2-6:2013  
EN ISO 15197:2015

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

EN IEC 63000:2018

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 2022-04-20

Name, function, signature, stamp: Daniel Kämmerer, Team Lead Regulatory Affairs

  
**Beurer GmbH**  
Söflinger Straße 218 • 89077 Ulm