


Attachment to Declaration of Conformity - HeartStart FRx

**Declaration of Conformity
Document Number/Rev.:** LC0285-101 Rev. AB

Description of the change: A minor design change was made to replace an obsolete component.
The new control indicator is product manufactured on or after
14Sep2022.


Assessment: This change has been assessed per MDCG 2020-3 guidance on significant
changes regarding the transitional provision under Article 120 of the
MDR with regard to devices covered by certificates according to MDD.
This change has been deemed to be non-significant.

Signature (signed for and on behalf of Philips Medical Systems)



Nadine Smith

Regulatory Affairs Program Manager – Emergency Care



Date of Issue: 13Oct2022

Place of Issue: Bothell, WA

Document #: LC0285-101-102

Date of Expiration: 26 May 2024